

PREBID CLARIFICATION

I-HUB QUANTUM TECHNOLOGY FOUNDATION, IISER, PUNE

First Floor, Main Academic Building, IISER Campus, Dr Homi Bhabha Rd, 411008
Tel: 020-25908647 Email: qtf_purchase@quantech.org.in Website: www.quantech.org.in

Refer I-HUB QTF's GeM tender no. GEM/2023/B/3754301 dated 05/08/2023 with file reference no. I-HUBQTF/PUR/23-24/008 for the procurement of Trinocular Upright Fluorescence Microscope. The Pre-Bid meeting was held on 16/08/2023.

At the outset, the Technical Committee welcomed all the Members and the Representatives of Prospective Bidders and briefed in general the scope of the tender and thereafter briefed the bidders on the salient features of the tender.

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of I-HUB QTF, IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents. Attached are the detailed technical and commercial queries with their clarifications (Annexure I).

The other terms & conditions of the notice issued on GeM and on I-HUB QTF's website (quantech.org.in) will remain unchanged. No more correspondence in this regard will be entertained. The meeting ended with a vote of thanks.

Sd/-

Project Director / Chief Executive Officer

TECHNICAL AND COMMERCIAL QUERIES AND CLARIFICATION – ANNEXURE I

Tendered Sr. No.	Query / Clarification Sought	Clarification / Amendment
-	<p>In the “Frame” section, it is requested to clarify as below:</p> <ul style="list-style-type: none"> • <i>We are not clear about Corresponded to RoHS regulations. Please explain.</i> 	<p>Tendered specification is amended from:</p> <p><i>“Microscope frame Built-in Köhler illumination for transmitted and reflected light, light present switch, light intensity manager switch, high color reproductivity LED light source, Lifetime of at least 50,000 Hours. Corresponded to RoHS regulations”</i></p> <p>to</p> <p><i>“Microscope frame Built-in Köhler illumination for transmitted and reflected light, light present switch, light intensity manager switch, high color reproductivity LED light source, Lifetime of at least 50,000 Hours”.</i></p>
-	<p>In the “Light Source” section, it is requested to amend the specifications as below:</p> <ul style="list-style-type: none"> • <i>Plases change 6 or more position fluorescence attachments.</i> 	<p>Tendered specification is amended from:</p> <p><i>“LED Light Source with 8 Position fluorescence attachment. Universal reflected light illuminator with field stop, aperture stop and 8-position mirror unit cassette, including UV protective shield, shield plate, indication plate and shield cap.”</i></p> <p>To</p> <p><i>“LED Light Source with 6 or more Position fluorescence attachment. Universal reflected light illuminator with field stop, aperture stop and 6 or more position mirror unit cassette, including UV protective shield, shield plate, indication plate and shield cap.”</i></p>
-	<p>In the “Camera and Software” section, it is requested to amend the specifications as below:</p> <ul style="list-style-type: none"> • <i>Cooled /non cooled Colour Camera.</i> • <i>C/F mount</i> 	<p>Tendered specification prevails.</p>